



ASAP INTERNATIONAL SDN. BHD. (1157980-X)

No. 1, Jalan Sitar 33/6, Seksyen 33, 40400 Shah Alam, Selangor, Malaysia.

Tel: +6 03 5191 0166

Fax: +6 03 5191 0702

Website: www.whyasap.com

Email: info@whyasap.com

PRODUCT SPECIFICATION

Product : ASAP Premium Nitrile Powder Free Examination Gloves

Product Description : Non-Sterile, Powder Free Nitrile Exam Gloves, Textured, Beaded, Blue Colour, Ambidextrous, Single Use Only.

Available Sizes : Extra Small, Small, Medium, Large, Extra Large

Packing Mode : 100 pcs per box, 10 boxes per carton

Quality Assurance : Meets EN 455 Standard

Test	Particular		Specification
Dimension (mm)	Palm Width	Extra Small	76 ± 3
		Small	84 ± 3
		Medium	94 ± 3
		Large	105 ± 3
		Extra Large	113 ± 3
	Length		Min. 240
	Thickness (Single wall)	Cuff (25 ± 5 from bead)	Min. 0.06
Palm (centre of palm)		Min. 0.08	
Finger (13 ± 3 from tip)		Min. 0.10	
Physical Properties	Force at Break (N)	Before Aging	Min. 6
		After Aging	Min. 6
Physical Appearance	Rolled beaded cuff Blue Colour Textured Surface Ambidextrous		

**Remark: The above data represent the norm of the product which may be changed without notice.

Handling and Storage

Store in a dry place at room temperature. Exposed product should be shielded from direct sunlight, intense artificial light, x-ray machines and other sources of ozone.

Caution

This product contains nitrile component which may cause an allergic reaction in some individuals. Do not expose these gloves to any person with a known or suspected sensitivity to nitrile. If allergic reaction occurs, discontinue use immediately and consult a physician. This product is not made with natural rubber latex.

Doc No.: PS AS2109-E (R1)

Rev Date: 18.10.2019

